

CERTIFICATE OF COMPLIANCE

BSL BIOSERVICE
Scientific Laboratories GmbH
Behringstr. 6/8
82152 Planegg
Germany

Phone +49 (0)89 899 650-0
Fax +49 (0)89 899 650-11



Testmaterial: EPDM 70 PEROXIDE CURED M534

Supplier: DUCI SRL
Via Monte Avaro, 69
24060 Chiuduno (BG)
Italy

Study performed: The following study was performed in order to determine the biocompatibility of the test item according to USP <87> "Biological Reactivity Tests, *in vitro* - Elution Test", ISO 10993-1: 2009, "Evaluation and testing within a risk management process", ISO 10993-5: 2009, "Tests for *in vitro* cytotoxicity" and ISO 10993-12: 2007, "Sample preparation and reference materials".

CYTOTOXICITY (BSL Project No. 123390)

Results: No leachable substances were released in cytotoxic concentrations from the test item. Hence, the test item meets the criteria of the USP as well as the ISO test for evaluation of cytotoxicity *in vitro*.

BSL BIOSERVICE Scientific Laboratories GmbH
Behringstraße 6/8
D-82152 Planegg
Germany

Handwritten signature in blue ink, appearing to read "B. Hoy".

Dr. Benjamin Hoy
Biological Safety Testing
Date: 02 May 2014